



ESTATION OF CONFORMIT

Certificate Nr: MDD-259

In conformance to the European Economic Commission 93/42/EEC Medical Devices Directive on harmonisation of laws, regulations and administrative documentation of Member States on Medical Devices and European Economic Commission directive 93/68/EEC amending Medical Devices Directive dated 22 July

the products manufactured For

CARINE EUROPE GMBH

Ammannstrasse No:12 86167 Augsburg / GERMANY

by the following manufacturer

AYDAĞ TEDAVİ ve SAĞLIK HİZMETLERİ SAN. ve TİC. A.S Orta Mah. Menderes Cad. No:7/3 K1 Bayrampaşa ISTANBUL / TURKEY

EN 14683:2019+AC:2019 Medical Face Masks

Technical standard EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods

Model: CRM-M01
Type IIR
are tested according to the following initial type tests by the manufacturer and ard EN 14683:2019+AC:2019 Medical face masks - Requirement of conformity the following initial type tests Blooming the following initial type tests by the manufacturer and the assessment of conformity the following initial type tests by the manufacturer and the following initial type tes For the assessment of conformity, the following documents were also applied to: Results of laboratory tests Ekoteks Laboratuvar Testing Laboratory Bacterial Filtration Efficiency, Microbial Cleanliness, Differential Pressure and Splash Resistance Pressure tests.

UNIVERSAL CERTIFICATION has evaluated production, design, intended use, risk evaluation according to safety purpose, product itself and add-on components (if exists) and product technical drawings of the medical face masks manufactured and designed for use during the medical operations or similar medical situations with same requirements which require restriction of infectious materials to be spread to patients. With this certificate, it is approved that the product fulfils all essential requirements and the related rules of 93/42/EEC Medical Devices Directive (MDD) Class I are applied. The information on the packaging for the above listed products covers the necessary information stated in Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex 1, §23, of the Medical Device Regulation (EU) 2017/745. This information includes; reference to EN 14683 standard, type of mask (as indicated in Table 1) and other relevant information given in EN ISO 15223-1:2016 and EN 1041:2008+A1:2013. It is considered to be suitable to attach a CE mark, as seen below, on your products in accordance with the information given in this certificate with publishing an EU Declaration of Conformity.

This certificate is issued on 11/09/2020 and valid until 10/09/2021 with the conditions that no change has been made with the product references and no change in the production process or not suspended or withdrawn for any reason.

ISTANBUL -11/09/2020

Medical & Personalized

Suat KAÇMAZ UNIVERSAL CERTIFICATION General Manager

Medical & Person



Medical & Persor

Verify the validity with the QR Code